

## Product specifications

---

Name	Anti-h SAA 2203 SPTN-5
Specificity	Antibody recognizes human serum amyloid A (SAA)
Description	Monoclonal mouse antibody, cultured in vitro under conditions free from animal-derived components
Product code	100289
Product buffer solution	50 mM Na-citrate, pH 6.0, 0.9 % NaCl, 0.095 % NaN <sub>3</sub> as a preservative
Shelf life and storage	18 months from manufacturing at 2–8 °C
Analyte description	Serum amyloid A, a family of apolipoproteins, is associated with high density lipoprotein during inflammatory states. The level of serum amyloid A (SAA) in the blood increases dramatically in response to tissue injury and inflammation. SAA also acts as a cytokine, influencing cell adhesion, migration, proliferation and aggregation. SAAs are implicated in several chronic inflammatory diseases, such as amyloidosis, atherosclerosis, and rheumatoid arthritis.

## Parameters tested on each lot

---

Product appearance	Liquid, may turn slightly opaque during storage
Product concentration	5.0 mg/ml (+/- 10 %)
Immunoreactivity	80–120 % compared to the reference sample in an FIA test
IEF Profile	6.8 – 7.7
Purity	≥ 95 %

## Parameters determined during product development

---

Subclass	IgG <sub>1</sub>
Association rate constant	1.1 x 10 <sup>5</sup> 1/Ms
Dissociation rate constant	3.0 x 10 <sup>-4</sup> 1/s
Affinity constant	K <sub>A</sub> = 3.5 x 10 <sup>8</sup> 1/M; K <sub>D</sub> = 2.8 x 10 <sup>-9</sup> M (= 2.8 nM)
Determination method	SPR analysis (ProteOn XPR36)
Determination antigen	SAA, Santa Cruz Biotechnology (Cat sc-4870, Lot A0308)



**Cross-reactivities** Recognizes human SAA1, human SAA2 and equine SAA.  
Does not recognize human SAA4, feline, canine or bovine SAA.

**Epitope** Conformational

**Pair recommendations**

		DETECTION						
		2201	2203	2205	2208	2211	2212	2213
CAPTURE	2201	-	+	+	-	+	-	-
	2203	+	-	+	-	+	-	+
	2205	-	-	-	-	+	+	-
	2208	-	-	-	-	-	+	-
	2211	+	-	-	-	-	-	+
	2212	-	-	+	+	-	-	-
	2213	+	-	-	-	-	-	-

Pairs 2203 (membrane) – 2201 (particles), 2201 – 2203, 2211 – 2213 and 2213 – 2201 are especially recommended for lateral flow (LF) assays.

Pairs 2201 – 2211, 2201 – 2203, 2203 – 2211, 2205 – 2211, 2201 – 2205 and 2203 – 2205 are especially recommended for immunoturbidimetric (IT) assays.

Pairs 2205 – 2212 and 2208 – 2212 are especially recommended for sandwich FIA assay with wide measuring range.

Please note that pair recommendations are based on results obtained by our laboratory. Equally good results may be obtained using other pairs and therefore these recommendations are only indicative.

**Platforms tested** FIA, IT, LF

**Antigens tested** Recognizes native human SAA, EastCoast Bio cat LA449, recombinant human SAA, Medix Biochemica cat 610070 and Lee Biosolutions cat 527-18.

TEMPERATURE, TIME	RESULT
-70 °C, 21 days	OK
-20 °C, 21 days	OK
+4 °C, 21 days	OK
+35 °C, 21 days	OK
+45 °C, 7 days	OK

Stability testing is performed in the product buffer to see whether different temperatures affect the antigen binding, charge or composition of the antibody. Please note that the shelf life given on the first page is based on real time stability testing at 2–8 °C in the product buffer

**Miscellaneous** -

**References** -

